

In claim 128, in line 2, the word "viewing" has been deleted and the word -- display -- inserted therefor.

In claim 128, in line 3, the word "demondulator" has been deleted and the word -- demodulator -- inserted therefor.

In claim 130, in line 2, the word "viewed" has been deleted and the word -- displayed -- inserted therefor.

In claim 132, in line 1, the word -- of -- has been inserted before the word "claim".

Claim 133 has been rewritten as follows:

133. A method of effecting medical screening and diagnosis, which comprises [effecting dual detection of] providing the apparatus of claim 92; positioning the apparatus on a subject; and detecting both stethoscopic and Doppler signals to effect said screening and diagnosis [using the apparatus of claim 92].

^{1-2 CE 3/15/08}
In claim 134, in lines ~~2-3~~, the phrase "wherein the apparatus is used to screen cardiovascular disease by measuring systolic pressure to establish" has been deleted and the phrase -- further comprising establishing -- has been inserted therefor.

In claim 135, in line 1, the phrase "wherein the screened cardiovascular disease is" has been deleted and the phrase -- further comprising determining whether the subject has -- has been inserted therefor.

4. The following is an examiner's statement of reasons for allowance: The prior art fails to show an apparatus for medical screening and diagnosis by dual detection of stethoscopic and Doppler signals, which comprises a housing which at least partially forms an ear trumpet provided with a membrane, wherein the housing is coupled to at